

# geolometer

R O C K   S O L I D L Y   R E L I A B L E



**GL-CT 110**

**Cement consistency  
condensation time meter**

# PRODUCT APPLICATION

GB/T1346-2011 "Cement net slurry standard consistency water consumption, condensation time, stability test method" is met by the cement consistency condensation time measuring device.

## Product Specifications

SPECIFICATIONS	GL-CT 110
Total Weight of Sliding Part	300 ± 1 g
Standard Consistency Test Rod	1 piece, 50 ± 1 mm length, $\Phi 10 \pm 0.05$ mm diameter
Initial Condensation Test Needles	2 pieces, 50 ± 1 mm length, $\Phi 1.13 \pm 0.05$ mm diameter
Final Condensation Test Needle	1 piece, 30 ± 1 mm length, $\Phi 1.13 \pm 0.05$ mm diameter
Dimensions (L x W x H)	170 × 110 × 300 mm
Weight	4.5 kg