

# geolometer

R O C K   S O L I D L Y   R E L I A B L E



## **GL-MI 30**

**Mixed mortar test parts standard  
maintenance box**

## PRODUCT APPLICATION

The strength and typical condensation time of the cement, concrete, and cement product specimens are tested using the standard maintenance box of the mixed mortar test sections.

## Product Specifications

SPECIFICATIONS	GL-MI 30
Temperature control instrument accuracy	20±1 °C
Humidity control	60–80% RH
Power supply voltage	AC220V ±10%, 50 Hz
Compressor power	138 W
Heater power	500 W
Studio size	580 × 500 × 1180 mm
Internal test frame equalizes 6-gauge size	580 × 490 × 180 mm
Dimensions	1000 × 700 × 1530 mm
Weight	130 kg